

Certificate of Conformance

Product: 10% Methane in Argon

Grade: Unanalyzed

The Methane/Argon cylinder(s) filled and distributed by Haun meet the following specifications based on our standard operation procedures.

Component	Specification
Methane	10%
Argon	Balance
Moisture/Dew Point °F	10.5 PPM/-76 °F

Note: Preparation tolerance of +/- 10% relative.